



## Declaration of Conformity

Manufacturer:	Authorized Representative:	Notified Body:
ResMed Pty. Ltd. 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia	ResMed SAS Parc Technologique de Lyon 292 Allée Jacques Monod 69791 Saint Priest Cedex France	TÜV SÜD Product Service GmbH Ridlerstraße 65 80339 München Germany

**Product:** AirFit F20 Non Magnetic

### Intended Use:

The AirFit F20 Non Magnetic is a non-invasive accessory used for channeling airflow (with or without supplemental oxygen) to a patient from a positive airway pressure (PAP) device such as a continuous positive airway pressure (CPAP) or bilevel system.

The AirFit F20 Non Magnetic is:

- to be used by patients weighing more than 30 kg for whom positive airway pressure therapy has been prescribed
- intended for single-patient reuse in the home environment and multipatient reuse in the hospital/institutional environment.

**Classification:** IIa according to Rule 2

**EMDN:** R0301010201 CPAP Masks

**Conformity Assessment Route:** Annex IX (excluding Chapter II), Regulation EU 2017/745

**Basic UDI-DI:** 619498EC1726N

**Common Specification:** N/A

We herewith declare that the above mentioned products are in conformity with the Council Regulation 2017/745 for medical devices.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer. This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

**EC Certificate Number:** G10 049861 0162 Rev. 02

**SRN:** AU-MF-000011753

Signed at Sydney, Australia on: 21 November 2023

DocuSigned by:

*Nicole Wilson*

F9F55744DEA64A0...

Nicole Wilson  
Person Responsible for Regulatory Compliance (PRRC)  
ResMed Pty. Ltd.

**EC172c.1**

First issued: 21 November 2023